

医用隔离面罩检验单

INSPECTION REPORT

编号:

产品名称 Name	医用隔离面罩 Medical Face Shield	规格 Specification	33cmX22cm
生产日期/MFG	2020-2-8	生产数量/QTY	100,000PCS
生产负责人/Production Manager		黄定海/Dinghai Huang	

检验内容 Inspection Content

编号 NO	检验项目 Item	具体要求 Specific Requirements	检验结果 Resule	不合格内容 Unqualified		
1	外观 Appearance	镜片表面应光洁、无明显划痕、杂质异物、晶点不得有破损或分离。 The surface of the lens should be smooth and clean, without obvious scratches, impurity for eign bodies, crystal points must not be damaged or separated.	合格 Qualified	NA		
		海绵表面完整无破损, 且与镜片粘和牢固 The surface of the sponge is intact and unbreakable is glued and secured with the lens	合格 Qualified	NA		
		橡胶筋与罩片连接牢固, 无破损 The rubber band is firmly attached to the shield without breakage	合格 Qualified	NA		
2	尺寸 Size	罩片全长/Length cm	33cm	±10%	合格 Qualified	NA
		罩片宽度/Width cm	22cm			
3	防雾性 Anti-fog	在垂直距离镜片任意面吹气 10 秒钟, 镜片表面不能有雾滴出现。 Blow on any surface of the lens at a vertical distance for 10 seconds, the surface of the lens must not have droplets.	合格 Qualified	NA		
检验结果 Result	产品合格 可以出货 Qualified products can be shipped					
质检员 Inspector	郑安娜 /Anna Zheng	质检负责人 Quality Manager	郑安娜 /Anna Zheng			

